

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 02/29/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:38
INTERNAL STANDARD	VERIFIED	10:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:41
BLANK TEST	0.00	10:42

