

SOLUTION CHANGE

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 04/03/2020

OPERATOR NAME:  
CHRISTOPHER R JORDAN  
Permit #: 46293  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 10:04 |
| INTERNAL STANDARD       | VERIFIED | 10:04 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 10:04 |
| BLANK TEST              | 0.00     | 10:05 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 10:05 |
| BLANK TEST              | 0.00     | 10:06 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 10:07 |
| BLANK TEST              | 0.00     | 10:07 |

