

SUPERVISOR TEST

STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108

Date: 04/18/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:08
INTERNAL STANDARD	VERIFIED	10:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12

