

SUPERVISOR TEST

-----  
STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108  
-----

Date: 02/10/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:25
INTERNAL STANDARD	VERIFIED	19:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:27
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:29

