

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 02/16/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:17
INTERNAL STANDARD	VERIFIED	11:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:17
BLANK TEST	0.00	11:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:19
BLANK TEST	0.00	11:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:20
BLANK TEST	0.00	11:21

