

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 02/03/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:14
INTERNAL STANDARD	VERIFIED	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18

