

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 02/29/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:38
INTERNAL STANDARD	VERIFIED	10:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:41
BLANK TEST	0.00	10:42

