

SOLUTION CHANGE

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 01/01/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:29
INTERNAL STANDARD	VERIFIED	08:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:31
BLANK TEST	0.00	08:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:33

