

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 01/16/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:15
INTERNAL STANDARD	VERIFIED	08:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:19

