

SUPERVISOR TEST

-----  
STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108  
-----

Date: 01/22/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:14
INTERNAL STANDARD	VERIFIED	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18

