

SUPERVISOR TEST

STONY POINT
TOWN POLICE DEPT.
DATAMASTER dmt: 112108

Date: 03/10/2020

OPERATOR NAME:
ROBERT BERTRAND
Permit #: 43319
Operator Agency/Dept: STONY POINT TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:43
INTERNAL STANDARD	VERIFIED	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:46

