

SUPERVISOR TEST

-----  
STONY POINT  
TOWN POLICE DEPT.  
DATAMASTER dmt: 112108  
-----

Date: 03/17/2020

OPERATOR NAME:  
ROBERT BERTRAND  
Permit #: 43319  
Operator Agency/Dept: STONY POINT TOWN PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:42
INTERNAL STANDARD	VERIFIED	18:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:43
BLANK TEST	0.00	18:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:45
BLANK TEST	0.00	18:46

